



**GW Allen Prover Inspection and Test Certification Package – GWP04**

Updated: 2/17/2025



## Gage Western Prover Inspection and Test Certification Package - GWP04

Updated: 2/17/2025

Prover Name: GW\_FMD\_025\_02\_P

Certification Date: 11/20/2024

Certification Due: 11/20/2027

Serial Number: 000737

Date of Fabrication: 11/2018

Max Working Pressure: 285 PSIG

Nominal Flow Rate: 2,500 GPM

### Package Contents:

1. Cover Sheet
2. Table of Contents
3. Primary Waterdraw
4. Secondary Waterdraw
5. Tertiary Waterdraw
6. Temperature Transmitter
7. Temperature Transmitter
8. Pressure Transmitter



## Prover Calibration Report

### Company Information

|           |                                        |                    |                  |
|-----------|----------------------------------------|--------------------|------------------|
| Customer: | GAGE WESTERN                           | Purchase Order #:  | 1                |
| Address:  | 7307 W INDUSTRIAL AVE MIDLAND TX 79706 | Work Order #:      | W/O-24-60690     |
| Owner:    | GAGE WESTERN                           | Calibrated Volume: | 20 GALLON BACKUP |

### Equipment Information

|                       |                      |                       |                |
|-----------------------|----------------------|-----------------------|----------------|
| Manufacturer:         | FMD                  | Calibration Date:     | 11/20/2024     |
| Model #:              | H44025(TT)1AASGAUVCS | Due Date:             | 11/20/2027     |
| Description:          | SVP                  | Temperature:          | 67 °F          |
| Serial #:             | 757                  | Humidity:             | 62 % RH        |
| ID #:                 | 0                    | Barometric Pressure:  | 30.240 inHg    |
| Tube Wall Thickness:  | 0.9655 inches        | Technician:           | ZACKARY CURRIE |
| Tube Inside Diameter: | 16.997 inches        | Procedure Used:       | API 4.9.4      |
| Conditions Statement  | USED, GOOD CONDITION | Calibration Location: | CPLM LAB       |

*This instrument listed on this certificate has been calibrated against standards traceable thru National Institute of Standards and Technology (NIST), to the International System of Units (SI). Coastal Flow Liquid Measurement's quality system complies with applicable requirements of ISO/IEC 17025:2017. The reported uncertainties and/or "test uncertainties ratios" (TUR's) are expressed as expanded uncertainty values at approximately the 95% confidence level using a coverage factor of K=2. Statement of compliance, where applicable, are based on test results falling within specified limits with the uncertainty of the measurement taken into account. All results contained herein relate only to item(s) calibrated. This certificate shall not be reproduced except in full, without the written permission of Coastal Flow Liquid Measurement.*

### Data Results

Volume of 75.7216 Liters at 60 F & 0 PSIG  
Volume of 20.0035 U.S. Gallons at 60 F & 0 PSIG  
Volume of 0.476274 Barrels (42 Gal) at 60 F & 0 PSIG



Customer Satisfaction Survey

Authorized By: Shannon Thomas, Waterdraw Supervisor  
Coastal Flow Liquid Measurement, Inc.

|                                    |              |                                             |                      |
|------------------------------------|--------------|---------------------------------------------|----------------------|
| Customer:                          | GAGE WESTERN | Prover Manufacturer:                        | FMD                  |
| Work Order #:                      | WO-24-60690  | Model #:                                    | H44025(TT)1AASGAAWCS |
| Service Location:                  | CFLM LAB     | Serial #:                                   | 737                  |
| Elevation:                         | 110          | Flow Tube Material:                         | 304 Stainless        |
| Standard Temperature (°F):         | 60.0         | Detector Bar Material:                      | 304 Stainless        |
| Standard Pressure (psig):          | 0.0          | Area Thermal Expansion Coefficient (Ga):    | 1.92E-05             |
| Reference Weight Density (gm/cc):  | 8.00         | Detector Thermal Expansion Coefficient (G): | 9.60E-06             |
| Field Test Weight Density (gm/cc): | 7.84         | Modulus of Elasticity (flow tube) (E):      | 2.80E+07             |
| Compressibility factor:            | 3.20E-06     | Flow Tube Inside Diameter (inches) (ID):    | 16.997               |
|                                    |              | Flow Tube Wall Thickness (inches) (WT):     | 0.966                |

|                                                      | Run 1     | Run 2                             | Run 3        | Run 4                                              | Run 5                                   |
|------------------------------------------------------|-----------|-----------------------------------|--------------|----------------------------------------------------|-----------------------------------------|
| Flow Rate (GPM)                                      | 17.643    | 13.950                            | 18.457       | 14.282                                             | 18.177                                  |
| Acquired Data                                        |           |                                   |              |                                                    |                                         |
| Expected Mass of Water (Ww) (gm)                     | 75496.6   | 75498.8                           | 75499.0      | 75499.2                                            | 75498.4                                 |
| Air Temperature (°F)                                 | 66.3      | 66.5                              | 66.9         | 66.9                                               | 67.1                                    |
| Detector Bar Temperature (Td) (°F)                   | 63.6      | 63.7                              | 63.8         | 63.9                                               | 63.9                                    |
| Prover Temperature (Tp) (°F)                         | 71.8      | 71.9                              | 72.0         | 72.0                                               | 72.1                                    |
| Prover pressure (Pp) (psig)                          | 30.0      | 30.0                              | 30.0         | 30.0                                               | 30.0                                    |
| Water Electrical Conductivity (µS)                   | 11.0      | 11.0                              | 11.0         | 11.0                                               | 11.0                                    |
| Calculated Densities                                 |           |                                   |              |                                                    |                                         |
| Density of Air (p <sub>a</sub> ) (gm/cc)             | 0.001204  | 0.001204                          | 0.001203     | 0.001203                                           | 0.001202                                |
| Density of Distilled Water (p <sub>w</sub> ) (gm/cc) | 0.997747  | 0.997734                          | 0.997721     | 0.997721                                           | 0.997708                                |
| Density of VSMOW (gm/cc)                             | N/A       | N/A                               | N/A          | N/A                                                | N/A                                     |
| IK value of VSMOW (µS/cm)                            | N/A       | N/A                               | N/A          | N/A                                                | N/A                                     |
| Density of Water in container (g/cm3)                | 0.997747  | 0.997734                          | 0.997721     | 0.997721                                           | 0.997708                                |
| Volume Calculation                                   |           |                                   |              |                                                    |                                         |
| Correction for Buoyancy of Weights (CBW)             | 1.001057  | 1.001058                          | 1.001057     | 1.001057                                           | 1.001056                                |
| Corrected Mass of Water (Mw) (gm)                    | 75576.400 | 75578.678                         | 75578.802    | 75579.003                                          | 75578.126                               |
| Volume of Water (Vw) (ml)                            | 75747.058 | 75750.328                         | 75751.440    | 75751.641                                          | 75751.749                               |
| Factors for Correction to Standard Conditions        |           |                                   |              |                                                    |                                         |
| Effect of Temp on Flow Tube Steel (CTS)              | 1.000261  | 1.000264                          | 1.000267     | 1.000268                                           | 1.000270                                |
| Effect of Press. On Flow Tube (CPSp)                 | 1.000019  | 1.000019                          | 1.000019     | 1.000019                                           | 1.000019                                |
| Effect of Press. on Water (CPLp)                     | 1.000096  | 1.000096                          | 1.000096     | 1.000096                                           | 1.000096                                |
| Temp. correction CTS for SVP (CTSp)                  | 0.999739  | 0.999736                          | 0.999733     | 0.999732                                           | 0.999730                                |
| WD Volume Corrected for CTSp                         | 75727.288 | 75730.330                         | 75731.214    | 75731.340                                          | 75731.296                               |
| WD value for single cal. Corrected CPSp, CPLp        | 75718.580 | 75721.622                         | 75722.506    | 75722.632                                          | 75722.588                               |
| Calibrated Prover Volume per Run, (CPV)              | 75718.580 | 75721.622                         | 75722.506    | 75722.632                                          | 75722.588                               |
| Calculated Deviation from Average Volume             | -0.0040%  | 0.0000%                           | 0.0012%      | 0.0014%                                            | 0.0013%                                 |
| Repeatability: 0.005%                                |           | 75721.586                         | ml           | 9.32                                               | Expanded Uncertainty by Unit of Measure |
| Environment Details                                  |           | 75.7216                           | liters       | 0.01                                               |                                         |
| Humidity: 62.0 % rh                                  |           | 0.0757216                         | cubic meters | 0.00                                               |                                         |
| Barometric Pressure: 30.2 in. Hg                     |           | 20.0035                           | gallons      | 0.00                                               |                                         |
| Ambient Temperature: 67.0 °F                         |           | 4620.8147                         | cubic inches | 0.57                                               |                                         |
| Notes:                                               |           | 0.476274                          | barrels      | 0.00                                               |                                         |
| NEW SEALS BEFORE WD                                  |           | Previous Calibration Date: 1/0/00 |              | 0.01% Expanded Uncertainty expressed as Percentage |                                         |
|                                                      |           | Previous Volume: 19.99514 gallons |              |                                                    |                                         |
|                                                      |           | % Change in BPV: 0.042 %          |              |                                                    |                                         |

Performed By:   
ZACKARY CURRIE

Witness: \_\_\_\_\_  
Witness: \_\_\_\_\_

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

For: GAGE WESTERN

Date: 11/20/2024

For:

Date: 11/20/2024

# CoastalFlow

by Quorum Software  
Gravimetric Calibration Method  
Equipment Traceability Report  
11/20/2024

Customer: GAGE WESTERN  
Report #: WAO-24-60890  
Scale Serial #: 43566166  
Scale Range (kg): 0 - 300 kg

Prover Model #: HH40025(111)1AASG/AWCS  
Prover Serial #: 7377  
Nominal Volume (gal): 19.9951  
Container Mass (kg): 35.000  
Expected Mass (kg): 75.650  
Target Mass (kg): 88.552 to 132.828

| Test Weights          |            |              |              |                |
|-----------------------|------------|--------------|--------------|----------------|
| Serial #              | Cert. Date | Nominal Mass | As Left Mass | Aggregate Mass |
| 43360                 |            | 100.0 kg     | -0.001004 kg | 99.9990 kg     |
| Total Aggregate Mass: |            |              |              | 99.9990 kg     |

| Verification | Target Mass (kg) | Scale Indication (kg) | Zero (kg) | Difference (kg) | Allowable Diff (kg) |
|--------------|------------------|-----------------------|-----------|-----------------|---------------------|
| Test 1       | 100.0000         | 100.0000              | 0.0004    | 0.0000          | 0.0050              |
| Test 2       | 100.0000         | 99.9998               | 0.0004    | 0.0002          | 0.0050              |
| Test 3       | 100.0000         | 100.0004              | 0.0000    | -0.0004         | 0.0050              |

| Equipment          | Serial #  | Range              | Calibration Date | Traceability # | Asset ID  |
|--------------------|-----------|--------------------|------------------|----------------|-----------|
| Temperature        | 456095    | 0-120 F            | 11/13/2024       | 92789          | 456095    |
| Pressure           | 574807    | 0-100 PSI          | 8/6/2024         | 92789          | 574807    |
| Scale 1            | 43566166  | 0 - 300 Kg         | 2/14/2023        | C04L           | 43566166  |
| Scale 2            | N/A       | N/A                | N/A              | N/A            | N/A       |
| Water Conductivity | 574506    | 0-10000 Ms         | 8/11/2022        | ND006067       | 574506    |
| Barometer          | 240348813 | 400 - 1070 mbh/hPa | 8/26/2024        | 400-14169058   | 230588623 |

Performed By:   
ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witness: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witness: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

Customer Name: GAGE WESTERN  
Service Location: CFLM LAB  
Work Order No: WO-24-60690

Manufacturer: FMD  
Prover Type: SVP  
Model No: H44025(TT)1AASGAWCS  
Serial No: 737  
Flow Tube Material: 304 SS  
Detector Bar Material: 304 SS  
Flow Tube Wall Thickness (WT): 0.9655 in  
Flow Tube Inside Diameter (ID): 16.997 in  
Previous Volume: 19.99514 gal  
Date: 12/6/2021

Elevation: 110 ft  
Relative Humidity: 62 % rh  
Barometric Pressure: 30.24 in. Hg  
Ambient Temperature: 67 °F

|                                    | Run 1          | Run 2          | Run 3          | Run 4          | Run 5          |
|------------------------------------|----------------|----------------|----------------|----------------|----------------|
| Run Time (sec)                     | <u>68</u>      | <u>86</u>      | <u>65</u>      | <u>84</u>      | <u>66</u>      |
| <b>Acquired Data</b>               |                |                |                |                |                |
| Scale #1 Mass of Water (Ww) (g)    | <u>75496.6</u> | <u>75498.8</u> | <u>75499.0</u> | <u>75499.2</u> | <u>75498.4</u> |
| Scale #2 Mass of Water (Ww) (g)    |                |                |                |                |                |
| Scale #3 Mass of Water (Ww) (g)    |                |                |                |                |                |
| Scale #4 Mass of Water (Ww) (g)    |                |                |                |                |                |
| Total Mass of Water (Ww) (g)       |                |                |                |                |                |
| Air Temperature (°F)               | <u>66.3</u>    | <u>66.5</u>    | <u>66.9</u>    | <u>66.9</u>    | <u>67.1</u>    |
| Detector Bar Temperature (Td) (°F) | <u>63.6</u>    | <u>63.7</u>    | <u>63.8</u>    | <u>63.9</u>    | <u>63.9</u>    |
| Prover Temperature (Tp) (°F)       | <u>71.8</u>    | <u>71.9</u>    | <u>72.0</u>    | <u>72.0</u>    | <u>72.1</u>    |
| Prover pressure (Pp) (psig)        | <u>30</u>      | <u>30</u>      | <u>30</u>      | <u>30</u>      | <u>30</u>      |
| Water Electrical Conductivity (µS) | <u>11</u>      | <u>11</u>      | <u>11</u>      | <u>11</u>      | <u>11</u>      |

Scale #1

| Zero in Kg   | Target Mass in Kg | Scale Indication in Kg |
|--------------|-------------------|------------------------|
| <u>.0004</u> |                   | <u>100.0000</u>        |
| <u>.0004</u> |                   | <u>99.9998</u>         |
| <u>.0000</u> |                   | <u>100.0004</u>        |

Scale #2

| Zero in Kg | Target Mass in Kg | Scale Indication in Kg |
|------------|-------------------|------------------------|
|            |                   |                        |
|            |                   |                        |
|            |                   |                        |

Performed By:   
ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witnessed By: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witnessed By: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

Comments: New Seals Before W/D



## Prover Calibration Report

### Company Information

Customer: GAGE WESTERN  
Address: 7907 W INDUSTRIAL AVE MIDLAND TX 79706  
Owner: GAGE WESTERN

Purchase Order #: 1  
Work Order #: WWO-24-60950  
Calibrated Volume: 15 GALLON

### Equipment Information

Manufacturer: FMD  
Model #: H40025(ITI)H40025GAWCS  
Description: SVP  
Serial #: 737  
ID #: 0  
Tube Wall Thickness: 0.0555 inches  
Tube Inside Diameter: 1.6997 inches

Calibration Date: 11/20/2024  
Due Date: 11/20/2027  
Temperature: 64 °F  
Humidity: 62 % RH  
Barometric Pressure: 30.240 inHg  
Technician: ZACHARY CURRIE  
Procedure Used: API 4.5.4  
Calibration Location: OPLANT LAE

Conditions Statement USED, GOOD CONDITION

*This instrument listed on this certificate has been calibrated against standards traceable thru National Institute of Standards and Technology (NIST), to the International System of Units (SI). Coastal Flow Liquid Measurement's quality system complies with applicable requirements of ISO/IEC 17025:2017. The reported uncertainties and or "test uncertainties ratios" (TUR's) are expressed as expanded uncertainty values at approximately the 95% confidence level using a coverage factor of K=2. Statement of compliance, where applicable, are based on test results falling within specified limits with the uncertainty of the measurement taken into account. All results contained herein relate only to item(s) calibrated. This certificate shall not be reproduced except in full, without the written permission of Coastal Flow Liquid Measurement.*

### Data Results

Volume of 56.8054 Liters at 60 F & 0 PSIG  
Volume of 15.0064 U.S. Gallons at 60 F & 0 PSIG  
Volume of 0.357295 Barrels (42 Gall) at 60 F & 0 PSIG



Customer Satisfaction Survey

Authorized By: Shannon Thomas, Waterdraw Supervisor  
Coastal Flow Liquid Measurement, Inc.

|                                    |              |                                              |                      |
|------------------------------------|--------------|----------------------------------------------|----------------------|
| Customer:                          | GAGE WESTERN | Prover Manufacturer:                         | FMD                  |
| Work Order #:                      | WO-24-60690  | Model #:                                     | H44025(TT)1AASGAAWCS |
| Service Location:                  | CFLM LAB     | Serial #:                                    | 737                  |
| Elevation:                         | 110          | Flow Tube Material:                          | 304 Stainless        |
| Standard Temperature (°F):         | 60.0         | Detector Bar Material:                       | 304 Stainless        |
| Standard Pressure (psig):          | 0.0          | Area Thermal Expansion Coefficient (Ga):     | 1.92E-05             |
| Reference Weight Density (gm/cc):  | 8.00         | Detector Thermal Expansion Coefficient (Gt): | 9.60E-06             |
| Field Test Weight Density (gm/cc): | 7.84         | Modulus of Elasticity (flow tube) (E):       | 2.80E+07             |
| Compressibility factor:            | 3.20E-06     | Flow Tube Inside Diameter (inches) (ID):     | 16.997               |
|                                    |              | Flow Tube Wall Thickness (inches) (WT):      | 0.966                |

|                                                       | Run 1     | Run 2                      | Run 3            | Run 4     | Run 5                                        |
|-------------------------------------------------------|-----------|----------------------------|------------------|-----------|----------------------------------------------|
| Flow Rate (GPM)                                       | 25.008    | 14.290                     | 20.937           | 13.437    | 21.958                                       |
| Acquired Data                                         |           |                            |                  |           |                                              |
| Expected Mass of Water (Ww) (gm)                      | 56639.0   | 56640.8                    | 56641.4          | 56643.2   | 56640.8                                      |
| Air Temperature (°F)                                  | 63.3      | 63.6                       | 63.7             | 63.9      | 64.2                                         |
| Detector Bar Temperature (Td) (°F)                    | 62.5      | 62.4                       | 62.3             | 62.3      | 62.4                                         |
| Prover Temperature (Tp) (°F)                          | 71.2      | 71.2                       | 71.3             | 71.3      | 71.4                                         |
| Prover pressure (Pp) (psig)                           | 30.0      | 30.0                       | 30.0             | 30.0      | 30.0                                         |
| Water Electrical Conductivity (µS)                    | 11.0      | 11.0                       | 11.0             | 11.0      | 11.0                                         |
| Calculated Densities                                  |           |                            |                  |           |                                              |
| Density of Air (ρ <sub>a</sub> ) (gm/cc)              | 0.001211  | 0.001210                   | 0.001210         | 0.001210  | 0.001209                                     |
| Density of Distilled Water (ρ <sub>st</sub> ) (gm/cc) | 0.997822  | 0.997822                   | 0.997810         | 0.997810  | 0.997797                                     |
| Density of VSMOW (gm/cc)                              | N/A       | N/A                        | N/A              | N/A       | N/A                                          |
| K value of VSMOW (µS/cm)                              | N/A       | N/A                        | N/A              | N/A       | N/A                                          |
| Density of Water in container (g/cm3)                 | 0.997822  | 0.997822                   | 0.997810         | 0.997810  | 0.997797                                     |
| Volume Calculation                                    |           |                            |                  |           |                                              |
| Correction for Buoyancy of Weights (CBW)              | 1.001064  | 1.001063                   | 1.001063         | 1.001063  | 1.001062                                     |
| Corrected Mass of Water (Mw) (gm)                     | 56699.264 | 56701.009                  | 56701.610        | 56703.412 | 56700.953                                    |
| Volume of Water (Vw) (ml)                             | 56823.025 | 56824.773                  | 56826.059        | 56827.865 | 56826.141                                    |
| Factors for Correction to Standard Conditions         |           |                            |                  |           |                                              |
| Effect of Temp on Flow Tube Steel (CTS)               | 1.000239  | 1.000238                   | 1.000239         | 1.000239  | 1.000242                                     |
| Effect of Press. On Flow Tube (CPSp)                  | 1.000019  | 1.000019                   | 1.000019         | 1.000019  | 1.000019                                     |
| Effect of Press. on Water (CPLp)                      | 1.000096  | 1.000096                   | 1.000096         | 1.000096  | 1.000096                                     |
| Temp. correction CTS for SVP (CTSp)                   | 0.999761  | 0.999762                   | 0.999761         | 0.999761  | 0.999758                                     |
| WD Volume Corrected for CTSp                          | 56809.444 | 56811.249                  | 56812.478        | 56814.283 | 56812.389                                    |
| WD value for single cal. Corrected CPSp, CPLp         | 56802.912 | 56804.716                  | 56805.945        | 56807.750 | 56805.856                                    |
| Calibrated Prover Volume per Run, (CPV)               | 56802.912 | 56804.716                  | 56805.945        | 56807.750 | 56805.856                                    |
| Calculated Deviation from Average Volume              | -0.0044%  | -0.0013%                   | 0.0009%          | 0.0041%   | 0.0007%                                      |
| Repeatability: 0.009%                                 |           | 56805.436                  | ml               | 8.89      | Expanded Uncertainty by Unit of Measure      |
| Environment Details                                   |           | 56.8054                    | liters           | 0.01      |                                              |
| Humidity: 62.0 % rh                                   |           | 0.0568054                  | cubic meters     | 0.00      |                                              |
| Barometric Pressure: 30.2 in. Hg                      |           | 15.0064                    | gallons          | 0.00      |                                              |
| Ambient Temperature: 64.0 °F                          |           | 3466.4804                  | cubic inches     | 0.54      |                                              |
| Notes:                                                |           | 0.357295                   | barrels          | 0.00      | Expanded Uncertainty expressed as Percentage |
| NEW SEALS BEFORE WD,<br>V2 SWITCH PREVIOUSLY CHANGED  |           | Previous Calibration Date: | 1/0/00           | 0.02%     |                                              |
|                                                       |           | Previous Volume:           | 15.00467 gallons |           |                                              |
|                                                       |           | % Change in BPV:           | 0.012 %          |           |                                              |

Performed By: ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witness: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witness: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

# CoastalFlow

by Quorum Software  
Gravimetric Calibration Method  
Equipment Traceability Report  
11/20/2024


Customer: GAGE WESTERN  
Report #: WD-14-6160  
Scale Serial #: 43366166  
Scale Range (kg): 0 - 300 kg

Prover Model #: #44025(TT)1A1SGAAW/C5  
Prover Serial #: 737  
Nominal Volume (gal): 15.0047  
Container Mass (kg): 35.000  
Expected Mass (kg): 56.799  
Target Mass (kg): 73.439 to 110.159

| Test Weights          |            |              |              |                |
|-----------------------|------------|--------------|--------------|----------------|
| Serial #              | Cert. Date | Nominal Mass | As Left Mass | Aggregate Mass |
| 43360                 |            | 100.0 kg     | -0.001004 kg | 99.9990 kg     |
| Total Aggregate Mass: |            |              |              | 99.9990 kg     |

| Verification | Target Mass (kg) | Scale Indication (kg) | Zero (kg) | Difference (kg) | Allowable Diff (kg) |
|--------------|------------------|-----------------------|-----------|-----------------|---------------------|
| Test 1       | 100.0000         | 100.0000              | 0.0004    | 0.0000          | 0.0050              |
| Test 2       | 100.0000         | 99.9998               | 0.0004    | 0.0002          | 0.0050              |
| Test 3       | 100.0000         | 100.0004              | 0.0000    | -0.0004         | 0.0050              |

| Equipment          | Serial #  | Range              | Calibration Date | Traceability # | Asset ID  |
|--------------------|-----------|--------------------|------------------|----------------|-----------|
| Temperature        | 456095    | 0-120 F            | 11/13/2024       | 92789          | 456095    |
| Pressure           | 574807    | 0-100 PSI          | 8/6/2024         | 92789          | 574807    |
| Scale 1            | 43566166  | 0 - 300 Kg         | 2/14/2023        | C04L           | 43566166  |
| Scale 2            | N/A       | N/A                | N/A              | N/A            | N/A       |
| Water Conductivity | 574506    | 0-10000 Ms         | 8/11/2022        | ND006067       | 574506    |
| Barometer          | 240348813 | 400 - 1070 mbh/hPa | 8/26/2024        | 400-14169058   | 230588623 |

Performed By:   
ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witness: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witness: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

Customer Name: GAGE WESTERN

Service Location: CFLM LAB

Work Order No: WO-24-60690

Manufacturer: FMD

Prover Type: SVP

Model No: H44025(TT)1AASGAAWCS

Serial No: 737

Flow Tube Material: 304 SS

Detector Bar Material: 304 SS

Flow Tube Wall Thickness (WT): 0.9655 in

Flow Tube Inside Diameter (ID): 16.997 in

Previous Volume: 15.00467 gal

Date: 12/6/2021

Elevation: 110 ft  
Relative Humidity: 62 % rh  
Barometric Pressure: 30.24 in. Hg  
Ambient Temperature: 64 °F

|                                    | Run 1   | Run 2   | Run 3   | Run 4   | Run 5   |
|------------------------------------|---------|---------|---------|---------|---------|
| Run Time (sec)                     | 36      | 63      | 43      | 62      | 41      |
| <b>Acquired Data</b>               |         |         |         |         |         |
| Scale #1 Mass of Water (Ww) (g)    | 56639.0 | 56640.8 | 56641.4 | 56643.2 | 56640.8 |
| Scale #2 Mass of Water (Ww) (g)    |         |         |         |         |         |
| Scale #3 Mass of Water (Ww) (g)    |         |         |         |         |         |
| Scale #4 Mass of Water (Ww) (g)    |         |         |         |         |         |
| Total Mass of Water (Ww) (g)       |         |         |         |         |         |
| Air Temperature (°F)               | 63.3    | 63.6    | 63.7    | 63.9    | 64.2    |
| Detector Bar Temperature (Td) (°F) | 62.5    | 62.4    | 62.3    | 62.3    | 62.4    |
| Prover Temperature (Tp) (°F)       | 71.2    | 71.2    | 71.3    | 71.3    | 71.4    |
| Prover pressure (Pp) (psig)        | 30      | 30      | 30      | 30      | 30      |
| Water Electrical Conductivity (µS) | 11      | 11      | 11      | 11      | 11      |

Scale #1

| Zero in Kg | Target Mass in Kg | Scale Indication in Kg |
|------------|-------------------|------------------------|
| • 0004     | 100.0             | 100.0000               |
| • 0004     | 100.0             | 99.9998                |
| • 0000     | 100.0             | 100.0004               |

Scale #2

| Zero in Kg | Target Mass in Kg | Scale Indication in Kg |
|------------|-------------------|------------------------|
|            |                   |                        |
|            |                   |                        |
|            |                   |                        |

Performed By:

  
ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witnessed By:

For: GAGE WESTERN

Date: 11/20/2024

Witnessed By:

For:

Date: 11/20/2024

Comments:

New Seals Before WD, V8 switch previously changed



## Prover Calibration Report

### Company Information

|           |                                        |                    |             |
|-----------|----------------------------------------|--------------------|-------------|
| Customer: | GAGE WESTERN                           | Purchase Order #:  | 1           |
| Address:  | 7907 W INDUSTRIAL AVE MIDLAND TX 79706 | Work Order #:      | WO-24-60650 |
| Owner:    | GAGE WESTERN                           | Calibrated Volume: | 5 GALLON    |

### Equipment Information

|                       |                      |                       |                |
|-----------------------|----------------------|-----------------------|----------------|
| Manufacturer:         | FMD                  | Calibration Date:     | 11/20/2024     |
| Model #:              | H14025(11)1AASGAAWCS | Due Date:             | 11/20/2027     |
| Description:          | SVP                  | Temperature:          | 66 °F          |
| Serial #:             | 737                  | Humidity:             | 62 % RH        |
| ID #:                 | 0                    | Barometric Pressure:  | 30.240 inHg    |
| Tube Wall Thickness:  | 0.9655 inches        | Technician:           | ZACKARY CURRIE |
| Tube Inside Diameter: | 16.997 inches        | Procedure Used:       | API 4.9.4      |
| Conditions Statement  | USED, GOOD CONDITION | Calibration Location: | CFM LAB        |

*This instrument listed on this certificate has been calibrated against standards traceable thru National Institute of Standards and Technology (NIST), to the International System of Units (SI). Coastal Flow Liquid Measurement's quality system complies with applicable requirements of ISO/IEC 17025:2017. The reported uncertainties and/or "test uncertainties ratios" (TUR's) are expressed as expanded uncertainty values at approximately the 95% confidence level using a coverage factor of K=2. Statement of compliance, where applicable, are based on test results falling within specified limits with the uncertainty of the measurement taken into account. All results contained herein relate only to item(s) calibrated. This certificate shall not be reproduced except in full, without the written permission of Coastal Flow Liquid Measurement.*

### Data Results

Volume of 18.9148 Liters at 60 F & 0 PSIG  
Volume of 4.99676 U.S. Gallons at 60 F & 0 PSIG  
Volume of 0.118970 Barrels (42 Gall) at 60 F & 0 PSIG



Customer Satisfaction Survey

Authorized By: Shannon Thomas, Waterdraw Supervisor  
Coastal Flow Liquid Measurement, Inc.

|                                    |              |                                              |                      |
|------------------------------------|--------------|----------------------------------------------|----------------------|
| Customer:                          | GAGE WESTERN | Prover Manufacturer:                         | FMD                  |
| Work Order #:                      | WO-24-60690  | Model #:                                     | H44025(TT)1AASGAAWCS |
| Service Location:                  | CFLM LAB     | Serial #:                                    | 737                  |
| Elevation:                         | 110          | Flow Tube Material:                          | 304 Stainless        |
| Standard Temperature (°F):         | 60.0         | Detector Bar Material:                       | 304 Stainless        |
| Standard Pressure (psig):          | 0.0          | Area Thermal Expansion Coefficient (Ga):     | 1.92E-05             |
| Reference Weight Density (gm/cc):  | 8.00         | Detector Thermal Expansion Coefficient (GI): | 9.60E-06             |
| Field Test Weight Density (gm/cc): | 7.84         | Modulus of Elasticity (flow tube) (E):       | 2.80E+07             |
| Compressibility factor:            | 3.20E-06     | Flow Tube Inside Diameter (inches) (ID):     | 16.997               |
|                                    |              | Flow Tube Wall Thickness (inches) (WT):      | 0.966                |

|                                                      | Run 1             | Run 2                             | Run 3        | Run 4     | Run 5                                        |
|------------------------------------------------------|-------------------|-----------------------------------|--------------|-----------|----------------------------------------------|
| Flow Rate (GPM)                                      | 11.090            | 8.555                             | 12.476       | 7.880     | 11.516                                       |
| Acquired Data                                        |                   |                                   |              |           |                                              |
| Expected Mass of Water (Ww) (gm)                     | 18859.2           | 18859.4                           | 18860.0      | 18859.6   | 18859.0                                      |
| Air Temperature (°F)                                 | 65.1              | 65.3                              | 65.6         | 65.7      | 65.8                                         |
| Detector Bar Temperature (Td) (°F)                   | 62.9              | 63.0                              | 63.1         | 63.2      | 63.3                                         |
| Prover Temperature (Tp) (°F)                         | 71.6              | 71.6                              | 71.7         | 71.7      | 71.8                                         |
| Prover pressure (Pp) (psig)                          | 30.0              | 30.0                              | 30.0         | 30.0      | 30.0                                         |
| Water Electrical Conductivity (µS)                   | 11.0              | 11.0                              | 11.0         | 11.0      | 11.0                                         |
| Calculated Densities                                 |                   |                                   |              |           |                                              |
| Density of Air (ρ <sub>a</sub> ) (gm/cc)             | 0.001207          | 0.001206                          | 0.001206     | 0.001206  | 0.001205                                     |
| Density of Distilled Water (ρ <sub>w</sub> ) (gm/cc) | 0.997772          | 0.997772                          | 0.997759     | 0.997759  | 0.997747                                     |
| Density of VSMOW (gm/cc)                             | N/A               | N/A                               | N/A          | N/A       | N/A                                          |
| K value of VSMOW (µS/cm)                             | N/A               | N/A                               | N/A          | N/A       | N/A                                          |
| Density of Water in container (g/cm3)                | 0.997772          | 0.997772                          | 0.997759     | 0.997759  | 0.997747                                     |
| Volume Calculation                                   |                   |                                   |              |           |                                              |
| Correction for Buoyancy of Weights (CBW)             | 1.001060          | 1.001059                          | 1.001059     | 1.001059  | 1.001058                                     |
| Corrected Mass of Water (Mw) (gm)                    | 18879.191         | 18879.372                         | 18879.973    | 18879.572 | 18878.953                                    |
| Volume of Water (Vw) (ml)                            | 18921.348         | 18921.529                         | 18922.378    | 18921.976 | 18921.583                                    |
| Factors for Correction to Standard Conditions        |                   |                                   |              |           |                                              |
| Effect of Temp on Flow Tube Steel (CTS)              | 1.000251          | 1.000252                          | 1.000254     | 1.000255  | 1.000258                                     |
| Effect of Press. On Flow Tube (CPSp)                 | 1.000019          | 1.000019                          | 1.000019     | 1.000019  | 1.000019                                     |
| Effect of Press. on Water (CPLp)                     | 1.000096          | 1.000096                          | 1.000096     | 1.000096  | 1.000096                                     |
| Temp. correction CTS for SVP (CTSp)                  | 0.999749          | 0.999748                          | 0.999746     | 0.999745  | 0.999742                                     |
| WD Volume Corrected for CTSp                         | 18916.599         | 18916.761                         | 18917.572    | 18917.151 | 18916.701                                    |
| WD value for single cal. Corrected CPSp, CPLp        | 18914.424         | 18914.586                         | 18915.397    | 18914.976 | 18914.526                                    |
| Calibrated Prover Volume per Run, (CPV)              | 18914.424         | 18914.586                         | 18915.397    | 18914.976 | 18914.526                                    |
| Calculated Deviation from Average Volume             | -0.0019%          | -0.0010%                          | 0.0033%      | 0.0010%   | -0.0014%                                     |
| Repeatability: 0.005%                                |                   | 18914.782                         | ml           | 23.45     | Expanded Uncertainty by Unit of Measure      |
| Environment Details                                  |                   | 18.9148                           | liters       | 0.02      |                                              |
| Humidity: 62.0 % rh                                  | Base Prover Vol.: | 0.0189148                         | cubic meters | 0.00      |                                              |
| Barometric Pressure: 30.2 in. Hg                     |                   | 4.99676                           | gallons      | 0.01      |                                              |
| Ambient Temperature: 66.0 °F                         |                   | 1154.2508                         | cubic inches | 1.43      |                                              |
|                                                      |                   | 0.118970                          | barrels      | 0.00      |                                              |
| Notes:                                               |                   | Previous Calibration Date: 1/0/00 |              | 0.12%     | Expanded Uncertainty expressed as Percentage |
| NEW SEALS BEFORE WD                                  |                   | Previous Volume: 4.99046 gallons  |              |           |                                              |
| V2 SWITCH PREVIOUSLY CHANGED                         |                   | % Change in BPV: 0.126 %          |              |           |                                              |

Performed By: LACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witness: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witness: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

Customer: GAGE WESTERN  
Report #: WAD-24-60890  
Scale Serial #: 43566166  
Scale Range (kg): 0 - 300 kg

Prover Model #: HANZE (IT) 1145640005  
Prover Serial #: 7377  
Nominal Volume (gal): 4.99005  
Container Mass (kg): 35.0000  
Expected Mass (kg): 118.891  
Target Mass (kg): 43.113 to 64.669

| Test Weights          |            |              |              |                |
|-----------------------|------------|--------------|--------------|----------------|
| Serial #              | Cert. Date | Nominal Mass | As Left Mass | Aggregate Mass |
| 43362                 |            | 50.0 kg      | 0.000546 kg  | 50.0005 kg     |
| Total Aggregate Mass: |            |              |              | 50.0005 kg     |

| Verification | Target Mass (kg) | Scale Indication (kg) | Zero (kg) | Difference (kg) | Allowable Diff (kg) |
|--------------|------------------|-----------------------|-----------|-----------------|---------------------|
| Test 1       | 50.0000          | 50.0000               | 0.0006    | 0.0000          | 0.0025              |
| Test 2       | 50.0000          | 50.0002               | 0.0000    | -0.0002         | 0.0025              |
| Test 3       | 50.0000          | 49.9998               | 0.0008    | 0.0002          | 0.0025              |

| Equipment          | Serial #  | Range              | Calibration Date | Traceability # | Asset ID  |
|--------------------|-----------|--------------------|------------------|----------------|-----------|
| Temperature        | 456095    | 0-120 F            | 11/13/2024       | 92789          | 456095    |
| Pressure           | 574807    | 0-100 PSI          | 8/6/2024         | 92789          | 574807    |
| Scale 1            | 43566166  | 0 - 300 Kg         | 2/14/2023        | C04L           | 43566166  |
| Scale 2            | N/A       | N/A                | N/A              | N/A            | N/A       |
| Water Conductivity | 574506    | 0-10000 Ms         | 8/11/2022        | ND006067       | 574506    |
| Barometer          | 240348813 | 400 - 1070 mbh/hPa | 8/26/2024        | 400-14169058   | 230588623 |

Performed By:   
ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witness: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witness: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

Customer Name: GAGE WESTERN  
Service Location: CFLM LAB  
Work Order No: WO-24-60690

Manufacturer: FMD  
Prover Type: SVP  
Model No: H44025(TT)1AASGAAWCS  
Serial No: 737  
Flow Tube Material: 304 SS  
Detector Bar Material: 304 SS  
Flow Tube Wall Thickness (WT): 0.9655 in  
Flow Tube Inside Diameter (ID): 16.997 in  
Previous Volume: 4.99046 gal  
Date: 12/6/2021

Elevation: 110 ft  
Relative Humidity: 62 % rh  
Barometric Pressure: 30.24 in. Hg  
Ambient Temperature: 66 °F

|                                    | Run 1          | Run 2          | Run 3          | Run 4          | Run 5          |
|------------------------------------|----------------|----------------|----------------|----------------|----------------|
| Run Time (sec)                     | <u>27</u>      | <u>35</u>      | <u>24</u>      | <u>38</u>      | <u>26</u>      |
| <b>Acquired Data</b>               |                |                |                |                |                |
| Scale #1 Mass of Water (Ww) (g)    | <u>18859.2</u> | <u>18859.4</u> | <u>18860.0</u> | <u>18859.6</u> | <u>18859.0</u> |
| Scale #2 Mass of Water (Ww) (g)    |                |                |                |                |                |
| Scale #3 Mass of Water (Ww) (g)    |                |                |                |                |                |
| Scale #4 Mass of Water (Ww) (g)    |                |                |                |                |                |
| Total Mass of Water (Ww) (g)       |                |                |                |                |                |
| Air Temperature (°F)               | <u>65.1</u>    | <u>65.3</u>    | <u>65.6</u>    | <u>65.7</u>    | <u>65.8</u>    |
| Detector Bar Temperature (Td) (°F) | <u>62.9</u>    | <u>63.0</u>    | <u>63.1</u>    | <u>63.2</u>    | <u>63.3</u>    |
| Prover Temperature (Tp) (°F)       | <u>71.6</u>    | <u>71.6</u>    | <u>71.7</u>    | <u>71.7</u>    | <u>71.8</u>    |
| Prover pressure (Pp) (psig)        | <u>30</u>      | <u>30</u>      | <u>30</u>      | <u>30</u>      | <u>30</u>      |
| Water Electrical Conductivity (µS) | <u>11</u>      | <u>11</u>      | <u>11</u>      | <u>11</u>      | <u>11</u>      |

Scale #1

| Zero in Kg    | Target Mass in Kg | Scale Indication in Kg |
|---------------|-------------------|------------------------|
| <u>0.0006</u> | <u>50</u>         | <u>50.0000</u>         |
| <u>0.0000</u> | <u>50</u>         | <u>50.0002</u>         |
| <u>0.0008</u> | <u>50</u>         | <u>49.9998</u>         |

Scale #2

| Zero in Kg | Target Mass in Kg | Scale Indication in Kg |
|------------|-------------------|------------------------|
|            |                   |                        |
|            |                   |                        |
|            |                   |                        |

Performed By: ZACKARY CURRIE

For: Coastal Flow Liquid Msmt.

Date: 11/20/2024

Witnessed By: \_\_\_\_\_

For: GAGE WESTERN

Date: 11/20/2024

Witnessed By: \_\_\_\_\_

For: \_\_\_\_\_

Date: 11/20/2024

Comments: V2 Switch Had Been Changed Previously  
new seals before W.D

25 September, 2018

Emerson Automation Solutions  
6021 Innovation Blvd  
Shakopee, MN 55379



## Calibration Data Sheet Consistent with ISO 10474 3.1 or EN 10204 3.1

|                                                                                                                                                                                                                  |                                                                                                                                                                                                                                                                                                      |
|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Customer Information</b><br><br>Name: FLOW MANAGEMENT DEVICES LLC<br>PO: 026932                                                                                                                               | <b>Manufacturer Information</b><br><br>Sales Order: 5333246<br>Line: 1                                                                                                                                                                                                                               |
| <b>Device Information</b><br><br>Device Type: Pressure Transmitter<br>Customer Ref/ID: 000-104178-SEN<br>Serial No: 0023974<br><br>Model No: 3051S2TG4A2E11A2EKDM8Q4<br><br>Output: Linear<br>Device ID: 2188257 | <b>Calibration Information</b><br><br>Factory: Shakopee, MN, USA<br>Station Name: SHAK_PRESSURE_CALIBRATION_05<br>Operator ID: 35588<br>Calibration Date: 09/24/2018 07:12:09 PM<br><br>Internal Ref #: 40106988  |
| <b>Permanent Tag Information</b>                                                                                                                                                                                 | <b>Wire-On Tag Information</b>                                                                                                                                                                                                                                                                       |

## Equipment Used

| EqNumber: | EqName:                     | CalDueDate:         |
|-----------|-----------------------------|---------------------|
| E3-58706  | Load Box                    | 3/20/2019 8:20:00AM |
| E3-60784  | Precision Resistance Decade | 2/9/2019 6:44:00AM  |
| P3-54297  | Pressure Controller         | 11/4/2018 2:30:00PM |
| P3-54940  | Pressure Controller         | 11/4/2018 3:05:00PM |

**Calibration Data**

Range: 0.000 TO 4,000.000 PSI

Reference of Calibration: Gauge  
Damping: 0.40 seconds

| % of Range | Applied Pressure | Requested Applied Pressure | Analog Output (mA) | % Span Error | Pass/Fail |
|------------|------------------|----------------------------|--------------------|--------------|-----------|
| 100.000    | 4000.000 PSI     | 4000.0000 PSI              | 20.00046           | 0.00288      | PASS      |
| 80.000     | 3200.000 PSI     | 3200.0000 PSI              | 16.80108           | 0.00675      | PASS      |
| 60.000     | 2400.000 PSI     | 2400.0000 PSI              | 13.59546           | -0.02837     | PASS      |
| 40.000     | 1600.000 PSI     | 1600.0000 PSI              | 10.39992           | -0.00050     | PASS      |
| 20.000     | 800.000 PSI      | 800.0000 PSI               | 7.20074            | 0.00463      | PASS      |
| 0.000      | 0.000 PSI        | 0.0000 PSI                 | 4.00064            | 0.00400      | PASS      |

This is to certify that the listed product meets the applicable Rosemount Specifications. Measuring and test equipment used in the manufacture and inspection of the listed product are traceable to the National Institute of Standards and Technology. The calibration system was designed to meet the intent of ANSI Z540-1-1994.



Heather Matthiesen  
Quality Manager

15 September, 2018

Emerson Automation Solutions  
6021 Innovation Blvd  
Shakopee, MN 55379



## Calibration Data Sheet Consistent with ISO 10474 3.1 or EN 10204 3.1

|                                                                                                                                                                                           |                                                                                                                                                                                                                                                                                 |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Customer Information</b><br><br>Name: FLOW MANAGEMENT DEVICES LLC<br>PO: 026716                                                                                                        | <b>Manufacturer Information</b><br><br>Sales Order: 5310382<br>Line: 1                                                                                                                                                                                                          |
| <b>Device Information</b><br><br>Device Type: Temperature Transmitter<br>Customer Ref/ID: 000-112012-SEN<br>Serial No: 0016073<br><br>Model No: 3144PD1A1KAM5T1C4Q4<br><br>Output: Linear | <b>Calibration Information</b><br><br>Factory: SHAKOPEE, MN, USA<br>Station Name: SHAK_TE<br>Operator ID: 73330<br>Calibration Date: 09/10/2018 05:58:22 AM<br><br>Internal Ref #: 40101475  |
| <b>Permanent Tag Information</b>                                                                                                                                                          | <b>Wire-On Tag Information</b>                                                                                                                                                                                                                                                  |

## Equipment Used

| EqNumber: | EqName:    | CalDueDate:           |
|-----------|------------|-----------------------|
| E3-60868  | Multimeter | 11/12/2018 11:06:00AM |

## Calibration Data

Sensor 1

Pt100 00385

4 Wire

-50.00 TO 150.00 DEG C

| % of Range | Measured Input | Digital Output | Analog Output | % Span Error | Pass/Fail |
|------------|----------------|----------------|---------------|--------------|-----------|
| 100.098    | 150.197 DEG C  | 150.201 DEG C  | 20.0161 mA    | 0.0021       | PASS      |
| 75.094     | 100.188 DEG C  | 100.185 DEG C  | 16.0151 mA    | 0.0004       | PASS      |
| 50.091     | 50.182 DEG C   | 50.176 DEG C   | 12.0138 mA    | -0.0047      | PASS      |
| 25.107     | 0.214 DEG C    | 0.213 DEG C    | 8.0175 mA     | 0.0024       | PASS      |
| 0.084      | -49.831 DEG C  | -49.829 DEG C  | 4.0136 mA     | 0.0005       | PASS      |

This is to certify that the listed product meets the applicable Rosemount Specifications. Measuring and test equipment used in the manufacture and inspection of the listed product are traceable to the National Institute of Standards and Technology. The calibration system was designed to meet the intent of ANSI Z540-1-1994.



Heather Matthiesen  
Quality Manager

23 October, 2018

Emerson Automation Solutions  
6021 Innovation Blvd  
Shakopee, MN 55379



## Calibration Data Sheet Consistent with ISO 10474 3.1 or EN 10204 3.1

|                                                                                                                                                                                           |                                                                                                                                                                                                                                                                                 |
|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| <b>Customer Information</b><br><br>Name: FLOW MANAGEMENT DEVICES LLC<br>PO: 027350                                                                                                        | <b>Manufacturer Information</b><br><br>Sales Order: 5378405<br>Line: 1                                                                                                                                                                                                          |
| <b>Device Information</b><br><br>Device Type: Temperature Transmitter<br>Customer Ref/ID: 000-112012-SEN<br>Serial No: 0019539<br><br>Model No: 3144PD1A1KAM5T1C4Q4<br><br>Output: Linear | <b>Calibration Information</b><br><br>Factory: SHAKOPEE, MN, USA<br>Station Name: SHAK_TE<br>Operator ID: 73432<br>Calibration Date: 10/23/2018 01:06:14 AM<br><br>Internal Ref #: 40119881  |
| <b>Permanent Tag Information</b>                                                                                                                                                          | <b>Wire-On Tag Information</b>                                                                                                                                                                                                                                                  |

## Equipment Used

| EqNumber: | EqName:    | CalDueDate:         |
|-----------|------------|---------------------|
| E3-60866  | Multimeter | 1/12/2019 7:46:00AM |

## Calibration Data

Sensor 1

Pt100 00385

4 Wire

-50.00 TO 150.00 DEG C

| % of Range | Measured Input | Digital Output | Analog Output | % Span Error | Pass/Fail |
|------------|----------------|----------------|---------------|--------------|-----------|
| 100.099    | 150.199 DEG C  | 150.200 DEG C  | 20.0161 mA    | 0.0011       | PASS      |
| 75.103     | 100.205 DEG C  | 100.212 DEG C  | 16.0156 mA    | -0.0050      | PASS      |
| 50.092     | 50.185 DEG C   | 50.183 DEG C   | 12.0133 mA    | -0.0094      | PASS      |
| 25.104     | 0.208 DEG C    | 0.205 DEG C    | 8.0157 mA     | -0.0059      | PASS      |
| 0.088      | -49.824 DEG C  | -49.828 DEG C  | 4.0133 mA     | -0.0049      | PASS      |

This is to certify that the listed product meets the applicable Rosemount Specifications. Measuring and test equipment used in the manufacture and inspection of the listed product are traceable to the National Institute of Standards and Technology. The calibration system was designed to meet the intent of ANSI Z540-1-1994.



Heather Matthiesen  
Quality Manager