

F417-70



14 Chippewa Road, Sherwood Park, Alberta T8A 3Y1
Ph: (780) 467-0972 | Fx: (780) 416-6913
Email: info@aircheklab.ca
Web: www.aircheklab.ca

Compressed Breathing Air Analysis

Ambient Pressure (Below 15 psig)

Test Date: December 5, 2024 Job # 82015
Sample Collection Date: December 5, 2024 Sample Received: December 5, 2024
Submitted By: Challand Pipelines
Compressor: Ingersol Rand XP185 Serial #: 406
Cylinder #: AMB 02 Unit #: .
Sampling Location: Challand Yard

Compounds Present	Maximum Allowable CSA Z180.1	Measured Quantity
ACM02		
Oxygen	20% - 22%	21 %
ACM05		
Nitrogen & Rare Gases	78% - 80%	79 %
Carbon Monoxide	5 ppm	< 0.75 ppm
Carbon Dioxide	600 ppm	452.8 ppm
Methane	10 ppm	3.0 ppm
Volatile Non-Methane Hydrocarbons	5 ppm	< 0.44 ppm
Volatile Halogenated Hydrocarbons	5 ppm	< 0.2 ppm
ACM04		
Odour	Free of Pronounced Odour	None
ACM03		
Oil, Particulates & Condensates	1 mg / m3	0.33 mg / m3

Sample Passed

Results apply to sample as received. Data provided during sample collection is used in the calculation of ACM03. Analysis was conducted under controlled environmental conditions.

This report may not be reproduced, except in full, without written consent from AirChekLab.

Val: 12-2

117472.0 R.1.2

Suggested next testing date in accordance
with CSA Standard

June 5, 2025

Analyst: G. McDougall

Entered by:

AirChekLab

Digitally signed by
Jeff McAlpine -
General Manager
Date: 2024.12.06
15:51:56 -07'00'